



## **DECLARATION OF CONFORMITY**

Product Identification:

Fluke 77 IV, 175, 177, 179 - Digital Multimeter

First declared: 2001

## Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO 9001:2015 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

Directive 2015/863/EU, UK Regulation 2016: RoHS

Directive 2014/30/EU, UK Regulation 2016: Electromagnetic Compatibility (EMC)

Directive 2014/35/EU, UK Regulation 2016: Low Voltage (LVD)

## Standards Used:

EN/BS 63000:2018 Assessment of electrical and electronic products for restriction of hazardous substances EN/BS 61326-1:2013 Electrical equipment for measurement, control and laboratory-EMC; General EN/BS 61326-2-2:2013 Particular EMC requirements for portable test, measuring and monitoring equipment EN/BS 61010-1:2010 Safety requirements for measurement, control and laboratory equipment; General EN/BS 61010-2-033:2012 Particular requirements for hand-held Multimeters for measuring Mains voltage

Special Conditions:

None

EU Importer/Distributor

Fluke Europe B.V. PO Box 1186 5602 BD Eindhoven The Netherlands

UK Importer/Distributor

Fluke Precision Measurement Ltd. 52 Hurricane Way Norwich, NR6 6JB United Kingdom

**Manufacturer** 

Fluke Corporation 6920 Seaway Boulevard M/S 143F Everett, WA 98203; USA

Henry Kuster

**EMEA Logistics and Compliance Manager** 

Fluke Europe B.V.

Thomas M. Smith P.E.

**Product Compliance Manager** Fluke Corporation